

Specifications Installation for the CM-9202-TM Calibration and Test unit for the CM-9202

SPECIFICATIONS

SIZE:	2.25" x 2.25" x 1"
WEIGHT:	2.5 oz. (nominal)
APPLIED IMPEDANCE:	Operator = 1.0 Meg (Impedance)
(Wrist Strap)	Bench = 6.8 Meg (Resistance)
ADJUSTMENTS:	No user adjustments necessary
CONTROLS:	Two Front Panel Push Button Switches

TEST PROCEDURE:

1. Plug the test box bench (Banana Male Plug) to common ground.
2. Connect the test box bench ground (Phono Male Plug) to the CM-9202 bench plug-in.
3. Plug the power supply as normal, to the CM-9202 power plug-in.
4. The CM-9202 bench led should be green.
5. Plug the test box operator (Banana Male Plug) into operator on the CM-9202.
6. Press the operator test box push button switch and the operator led on the CM-9202 should change from a blinking red to green.
7. Release the operator test box push button and the operator led on the CM-9202 should change to a blinking red, with momentary beeps up to a maximum of 7 to 12 beeps.
8. Press the bench test box push button and the bench led on the CM-9202 should change to red, with constant beeps.
9. Release the bench test box push button and the bench led on the CM-9202 should change to green, with no beeps
10. If the monitor did not pass as per the test procedure. Call our Technical Support Hot Line at (877) 463-9376.

CM-920TM INSTALLATION & TEST PROCEDURES

PRODUCT
CM-920TM

ITEM NUMBER
CM-920TM

DATA SHEET



1-877-GND-ZERO (463-9376)
FAX: 941-751-7586
4916 26TH Street West #100
Bradenton, Florida 34207
EMAIL: sales@gndzero.com
www.gndzero.com